

cal 17122013

### Meter Calibration Report

Station: Remote Oprtns Cntrlr  
Meter ID: 19028  
Meter Description: Molinos Azteca  
Operator: LOI

#### Differential Pressure Calibration (AIN A 1):

12-17-2013,13:44:23	Set Low Reading	AS Found:	0.000	Actual:	0.000	Difference:	0.0000%
12-17-2013,13:48:32	Set High Reading	AS Found:	62.000	Actual:	62.000	Difference:	0.0000%
12-17-2013,13:51:12	Set Mid Point 1	AS Found:	15.487	Actual:	15.500	Difference:	-0.0203%
12-17-2013,13:52:44	Set Mid Point 2	AS Found:	31.006	Actual:	31.000	Difference:	0.0101%
12-17-2013,13:54:39	Set Mid Point 3	AS Found:	46.493	Actual:	46.500	Difference:	-0.0118%

#### Temperature Calibration (AIN A 3):

12-17-2013,13:55:45	Set Low Reading	AS Found:	0.030	Actual:	0.000	Difference:	0.0758%
12-17-2013,14:06:20	Set High Reading	AS Found:	40.106	Actual:	40.000	Difference:	0.2653%
12-17-2013,14:16:21	Set Mid Point 1	AS Found:	10.019	Actual:	10.000	Difference:	0.0472%
12-17-2013,14:24:33	Set Mid Point 2	AS Found:	20.020	Actual:	20.000	Difference:	0.0504%
12-17-2013,14:32:26	Set Mid Point 3	AS Found:	29.985	Actual:	30.000	Difference:	-0.0378%

#### Static Pressure Calibration (AIN A 2):

12-17-2013,14:34:38	Set Low Reading	AS Found:	0.000	Actual:	0.000	Difference:	0.0000%
12-17-2013,14:38:17	Set High Reading	AS Found:	1499.629	Actual:	1500.000	Difference:	-0.0248%
12-17-2013,14:39:40	Set Mid Point 1	AS Found:	376.302	Actual:	375.000	Difference:	0.0868%
12-17-2013,14:40:18	Set Mid Point 2	AS Found:	750.062	Actual:	750.000	Difference:	0.0041%
12-17-2013,14:41:28	Set Mid Point 3	AS Found:	1125.836	Actual:	1125.000	Difference:	0.0558%